

# Material Specification

Declaration of Compliance with the Order EN 10204-2.1

## RetractoFit Bio

Ref 237440, 237480

Hamilton Bonaduz AG, Bonaduz, Switzerland

hereby declares that the wetted parts fulfill the following material specifications:

Wetted Parts	Material	Regulatory/Specification
Body	Stainless steel 1.4435	R <sub>a</sub> < 0.4 µm (N5)
Insertion Tube	Stainless steel 1.4435	R <sub>a</sub> < 0.4 µm (N5)
O-ring(s)	EPDM	FDA 21 CFR 177.2600 EU 1935/2004 EN ISO 10993-5 USP <87> USP <88> class VI (121 °C)

Hamilton Bonaduz AG



Dr. Jörg Pochert  
Vice President Process Analytics



Denise Ulke  
Quality Manager Process Analytics

Bonaduz, 2017 – 03-01

REF 695204/01

**HAMILTON** 

Web: [www.hamiltoncompany.com](http://www.hamiltoncompany.com)  
USA: 800-648-5950  
Europe: +41-58-610-10-10

**Hamilton Americas & Pacific Rim**  
4970 Energy Way  
Reno, Nevada 89502 USA  
Tel: +1-775-858-3000  
Fax: +1-775-856-7259  
[sales@hamiltoncompany.com](mailto:sales@hamiltoncompany.com)

**Hamilton Europe, Asia & Africa**  
Via Crusch 8  
CH-7402 Bonaduz, Switzerland  
Tel: +41-58-610-10-10  
Fax: +41-58-610-00-10  
[contact.pa.ch@hamilton.ch](mailto:contact.pa.ch@hamilton.ch)