

Declaration of Quality

CONDUCELL 4USF-PG-TL 225



Product Identification

Ref		237670	
SN		10696	
Lot		111196936	
Heat Number	Shaft	54881	
	Electrodes	118032	flat
		0290354	conical

Factory Test Conditions

Tests were performed according to	606182
Temperature	21.05 °C

Factory Test Results

	Test Specification	Result
Cell Constant in KCl @1350 μ S/cm	0.360 \pm 0.014 / cm	0.367 / cm
Quality Test successfully passed	2021-01-27	



This declaration is generated automatically and valid without signature.

REF 10069014/00