

Declaration of Quality

Incyte Arc - 120



Product Identification

Ref	243950-0211	
SN	1617	
Lot	111203548	
Heat Number	Electrodes	196611
	Shaft	54881
Firmware	CDCUM001	

Factory Test Conditions

Tests were performed according to	111000019
Temperature	23.79 °C

Factory Test Results

Permittivity in KCl @80 mS/cm	Passed
Permittivity in KCl @1 mS/cm	Passed
Cell constant in KCl @10 mS/cm	Passed
Quality Test successfully passed	2021-02-22



This declaration is generated automatically and valid without signature.

REF 10069014/00