

Declaration of Quality

OXYFERM FDA MS 225



Product Identification

| | |
|-------------|-------------|
| Ref | 237715 |
| SN | 4128 |
| Lot | 111178249 |
| Heat Number | Shaft 54881 |
| | Cap 849072 |

Factory Test Conditions

| | |
|-----------------------------------|-----------|
| Tests were performed according to | 111000085 |
| Temperature | 25.6 °C |

Factory Test Results

| | Test Specification | Result |
|----------------------------------|--------------------|----------|
| Actual current in air (25 °C) | 40 ... 80 nA | 52.85 nA |
| Quality Test successfully passed | 2020-11-16 | |



This declaration is generated automatically and valid without signature.

REF 10069014/00